



**广 微 测**  
Gmicro Testing

# GUANGDONG DETECTION CENTER OF MICROBIOLOGY

## REPORT FOR ANALYSIS

**Report №.**

2020FM24940R01E

**Name of Sample**

Air Purifying Disinfector

**Applicant**

Qierling (Beijing) Health Technology Co., Ltd.

**Test Type**

Entrustment Test



**Address:** Building 66, No.100 Central Xian Lie Road, Guangzhou, China

**Postcode:** 510070

**Tel:** +86 20 87137666

**Fax :** +86 20 87137668

**Website :** [www.gddcm.com](http://www.gddcm.com)




**GUANGDONG DETECTION CENTER OF MICROBIOLOGY**

**REPORT FOR ANALYSIS**



Report №.:2020FM24940R01E Verification Code: 79432168

Name of Sample	Air Purifiering Disinfectior	Test Type	Entrustment Test
Applicant	Qierling (Beijing) Health Technology Co., Ltd.	Address	No 101-42/101-43 (Dongsheng district), 9th Floor, No 1 Building, No 8th, Heiquan Road, Haidian District, Beijing City.
Sample Source	Submitted for Testing by the Applicant	Sample Quantity	1 complete machine + 3 sets of filters
Spec and Lot № of Sample	DS-X400N	State and Characteristic	Machine
Sample Received Date	2020-08-27	Test Completion Date	2020-09-27
Test Standard and Method	Refer to GB 21551.3-2010 Appendix A		
Item Tested	Air virus elimination effect		
Test Conclusion	<p>The test data of the sample(s) is attached to the page(s) of this report.</p> <p style="text-align: right;">Issue Date: 2020-12-08 (Official Seal)</p>		
Remarks	<p>Main inspection model: DS-X400N and models: DS-X400W DS-P400 are the same kind of products. (Communication module is different, others are the same) (provided by the applicant)</p>		



**Editor:** Wuqiuyan.

**Verifier:** Li Sujuan

**Approver:** Xie Xiaobao



GUANGDONG DETECTION CENTER OF MICROBIOLOGY

ANALYSIS AND TEST RESULT

Report No.: 2020FM24940R01E

Virus and host cell	Action Time	Serial Number	Air virus content (TCID <sub>50</sub> /m <sup>3</sup> )	Removal rate (%)
H3N2 Host cell: MDCK	0 (CK)	1	1.94×10 <sup>6</sup>	
		2	1.94×10 <sup>6</sup>	
		3	2.44×10 <sup>6</sup>	
	1h	1	<97.3	>99.99
		2	<97.3	>99.99
		3	<97.3	>99.99

(Blank below)



Remarks	The samples open the maximum wind speed and plasma functions in the 30m <sup>3</sup> chamber for test.
---------	--



## Notice Items

1. The Test report is invalid if not affixed with Authorized Stamp of Test and Paging Seal.
2. The Test report is invalid without signature of verifier and approver.
3. The Test report is invalid if being supplemented, deleted or altered.
4. Without prior written permission, the report cannot be reproduced, except in full.
5. Unless otherwise stated, the results shown in this test report refer only to the sample(s) submitted.
6. Any dispute of the report must be raised to the testing body within 15 days after the report is received, exceeding which the dispute will not be accepted。
7. For the tested sample(s) submitted by the applicant, the sample information in the test report is provided by the applicant and the laboratory is not responsible for its authenticity.
8. This test report is for reference only to the applicant and does not have a proof of effect for others.

