



201819000873

TEST REPORT



中国认可
国际互认
检测
TESTING
CNAS L7673

Applicant : Qierling (Beijing) Health Technology Co., Ltd.
Address : No. 101-42/101-43 (Dongsheng district), 9th Floor, No 1 Building, No 8th , Heiquan Road, Haidian District, Beijing City.

The following merchandise was (were) submitted and identified by the client as:

Name of Sample : 720 DS-S800 Air Purifying Disinfectant
Test Type : Commission
Sample Quantity : 1
Model : DS-S800
Batch No. : /
Brand : 720
Manufacturer: Healthlead Corporation Limited
Sample Received : 2020/07/15
Test Period : 2020/07/15-2020/08/20
Test Items : Please refer to next page(s).
Test Method : Please refer to next page(s).
Test Result : Please refer to next page(s).
Sample Description : Machine
Note: /

Edited by: 黄婉盈

Approved by: 叶智坚

Checked by: 叶智坚

Official Seal: 广州中科检测技术服务有限公司 检验检测专用章

TEST RESULTS (1):

Table 1 Summary of test results							
Chapter	Test Item(s)		Unit	Test Result(s)		Limiting Value	Test Method(s)
4.2.1	Removal rate (simulated field test)	<i>Staphylococcus aureus</i>	%	60min	>99.99	≥50%	GB 21551.3-2010
		<i>Escherichia coli</i>		60min	>99.99		
		<i>Candida albicans</i>		60min	>99.99		
		<i>Aspergillus niger</i>		60min	>99.99		

***** TO BE CONTINUED *****

检测



TEST RESULTS (2):

Table 2 Test data of removal rate (simulated field test)								
Test bacteria	Test time (min)	Test number	Control group			Test group		Removal rate K_t (%)
			Colony count before test V_0 (CFU/m ³)	Colony count after test V_t (CFU/m ³)	Natural decay rate N_t (%)	Colony count before test V_1 (CFU/m ³)	Colony count after test V_2 (CFU/m ³)	
<i>Staphylococcus aureus</i>	60	1	7.18×10 ⁴	5.49×10 ⁴	23.54	7.38×10 ⁴	<6	>99.99
		2	9.01×10 ⁴	6.83×10 ⁴	24.20	9.21×10 ⁴	<6	>99.99
		3	8.55×10 ⁴	6.70×10 ⁴	21.64	8.84×10 ⁴	<6	>99.99
		Mean						
<i>Escherichia coli</i>	60	1	7.50×10 ⁴	5.17×10 ⁴	31.07	7.72×10 ⁴	<6	>99.99
		2	6.00×10 ⁴	4.37×10 ⁴	27.17	6.14×10 ⁴	<6	>99.99
		3	6.50×10 ⁴	4.65×10 ⁴	28.46	6.74×10 ⁴	<6	>99.99
		Mean						
<i>Candida albicans</i>	60	1	9.14×10 ⁴	6.72×10 ⁴	26.48	8.81×10 ⁴	<6	>99.99
		2	8.46×10 ⁴	6.55×10 ⁴	22.58	8.17×10 ⁴	<6	>99.99
		3	7.10×10 ⁴	5.59×10 ⁴	21.27	7.29×10 ⁴	<6	>99.99
		Mean						
<i>Aspergillus niger</i>	60	1	5.70×10 ⁴	4.16×10 ⁴	27.02	5.54×10 ⁴	<6	>99.99
		2	6.32×10 ⁴	4.68×10 ⁴	25.95	6.01×10 ⁴	<6	>99.99
		3	6.67×10 ⁴	4.67×10 ⁴	29.99	6.47×10 ⁴	<6	>99.99
		Mean						

***** TO BE CONTINUED *****

Inspection instructions:

1. Test method

GB 21551.3-2010 Antibacterial and cleaning function for household and similar electrical appliances-Particular requirement of air cleaner (Annex A)

2. Test microorganism

Staphylococcus aureus ATCC6538, *Escherichia coli* 8099, *Candida albicans* ATCC10231, *Aspergillus niger* CMCC98003

3. Test conditions

- 1) Environment temperature:(20~25)°C
- 2) Environment humidity:(50~70)%RH

4. Test equipment

Test chamber (30 m³), six-stage sieve sampler (FA-1), Microbial aerosol generator, NA, PDA, SDA.

5. Operation conditions of the machine

Set the switch to position "Maximum Wind Speed".

6. Computational formula

$$\text{Natural decay rate } N_t(\%) = \frac{V_0 - V_t}{V_0} \times 100$$

where: V_0 = Colony count before test of control group; V_t = Colony count after test of control group

$$\text{Removal Rate } K_t(\%) = \frac{V_1 \times (1 - N_t) - V_2}{V_1 \times (1 - N_t)} \times 100$$

where: V_1 = Colony count before test of test group; V_2 = Colony count after test of test group.

***** TO BE CONTINUED *****

SAMPLE PHOTO



***** END OF REPORT *****

用章

Statement

1. This report is issued by The Guangzhou CAS Test Technical Services Co., Ltd (hereinafter referred to as "Our Company").
2. This report is invalid if not affixed with authorized stamp of test and paging seal.
3. This report is invalid without signature of verifier and approver.
4. This report is invalid if being supplemented, deleted or altered.
5. Without written permission of our Company, this report can not be reproduced in part (except in whole).
6. The result(s) shown in this report refer only to the sample(s) tested, but do not apply to the same batch, size or brand of products (except the test samples) or to prove the related methods of making, processing or production of the test sample(s), or the correctness and rationality of processes or process.
7. Objections to this report must be submitted to our Company within 15 days. Otherwise, it will automatically deem to have accepted this report.
8. The Client shall be responsible for the accuracy, authenticity and completeness of the samples and information submitted for inspection, and the disputes arising therefrom shall be borne by the Client.
9. As any reports is issued as a result of this application for testing services, our Company will strictly keep confidentiality to the Clients. Except where disclosure is required on the basis of laws, regulations, judgments, and rulings (including in accordance with summons, court, or government proceedings).
10. The result(s) or conclusion(s) shown in this report about the description of the characteristics, composition, properties or quality are based on the specific time, methods and applicable criteria. Using different methods and criteria or under different environmental conditions for testing may come to different conclusions.
11. Since our Company's causes lead to modify the contents of this report, our Company shall reissue this report and bear the modification cost. The Client shall return the original report. Since the Client's causes lead to modify the contents of this report, the Client need to submit an application form for the change of report to our Company. The Client shall bear the modification cost and return the original report if our Company approves to reissue this report.
12. The English version of this statement is translated from the Chinese one. If there is any disagreement between them, the Chinese version will be the final explanation.